

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
 Data File : VX035040.D
 Acq On : 12 Apr 2023 12:39
 Operator : JC/MD
 Sample : 02288-02DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GROUND-WATERDL

Quant Time: Apr 13 02:32:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	270156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	462690	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	416990	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	187975	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	189798	45.938	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.880%
35) Dibromofluoromethane	5.385	113	147663	48.251	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	8.647	98	554409	49.081	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.160%
62) 4-Bromofluorobenzene	11.079	95	209474	47.921	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.840%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	22192	13.222	ug/l	97
20) Methylene Chloride	2.794	84	2176	0.657	ug/l	93
24) 1,1-Dichloroethane	3.605	63	4146	0.703	ug/l	98
27) cis-1,2-Dichloroethene	4.483	96	10385	2.898	ug/l	91
32) 1,1,1-Trichloroethane	5.379	97	6610	1.293	ug/l #	51
37) Ethyl Acetate	4.721	43	14680	3.097	ug/l	99
44) Trichloroethene	7.129	130	7868	2.368	ug/l	99
52) Toluene	8.720	92	2795	0.329	ug/l	98
64) Tetrachloroethene	9.275	164	46507	15.585	ug/l	97
68) m/p-Xylenes	10.299	106	9883	1.646	ug/l	99
69) o-Xylene	10.640	106	5851	0.995	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	2844	0.243	ug/l	100
94) Hexachlorobutadiene	13.725	225	621	0.401	ug/l	70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
Data File : VX035040.D
Acq On : 12 Apr 2023 12:39
Operator : JC/MD
Sample : 02288-02DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GROUND-WATERDL

Quant Time: Apr 13 02:32:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

