

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
 Data File : VX035043.D
 Acq On : 12 Apr 2023 13:49
 Operator : JC/MD
 Sample : 02293-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IDW-LIQ-FTWZ1A-041123

Quant Time: Apr 13 02:34:22 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	261009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	445212	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401889	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	188970	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	180929	45.326	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.660%
35) Dibromofluoromethane	5.391	113	137554	46.712	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.420%
50) Toluene-d8	8.647	98	534172	49.146	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	203929	48.484	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.960%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	9361	5.773	ug/l	99
27) cis-1,2-Dichloroethene	4.495	96	2715	0.784	ug/l	78
44) Trichloroethene	7.129	130	1488	0.465	ug/l	92
52) Toluene	8.720	92	28087	3.440	ug/l	100
68) m/p-Xylenes	10.305	106	3834	0.663	ug/l	85
69) o-Xylene	10.640	106	1354	0.239	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	5159	0.438	ug/l	97
89) n-Butylbenzene	12.335	91	2954	0.307	ug/l	90
93) 1,2,4-Trichlorobenzene	13.591	180	1095	0.309	ug/l	95
94) Hexachlorobutadiene	13.725	225	1022	0.657	ug/l	95
95) Naphthalene	13.780	128	4896	0.389	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
Data File : VX035043.D
Acq On : 12 Apr 2023 13:49
Operator : JC/MD
Sample : 02293-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IDW-LIQ-FTWZ1A-041123

Quant Time: Apr 13 02:34:22 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

