

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
 Data File : VX035048.D
 Acq On : 12 Apr 2023 16:12
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 13 02:36:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	285492	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	484687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	439475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	206404	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	190568	43.646	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	87.300%
35) Dibromofluoromethane	5.385	113	145614	45.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.840%
50) Toluene-d8	8.647	98	578911	48.924	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.079	95	221264	48.321	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.640%
Target Compounds						
73) Isopropylbenzene	10.963	105	81439	5.386	ug/l	100
85) sec-Butylbenzene	11.890	105	5487	0.380	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
Data File : VX035048.D
Acq On : 12 Apr 2023 16:12
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 13 02:36:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

