

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
 Data File : VX035051.D
 Acq On : 12 Apr 2023 17:24
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 13 02:38:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	278904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	466342	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	426300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	209217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	183035	42.911	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.820%
35) Dibromofluoromethane	5.385	113	137195	44.479	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.960%
50) Toluene-d8	8.646	98	562494	49.407	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.820%
62) 4-Bromofluorobenzene	11.079	95	220360	50.017	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.040%
Target Compounds						
					Qvalue	
73) Isopropylbenzene	10.963	105	17554	1.145	ug/l	99
84) 1,2,4-Trimethylbenzene	11.756	105	5731	0.440	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041323\
Data File : VX035051.D
Acq On : 12 Apr 2023 17:24
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 13 02:38:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

