

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041422\
 Data File : VX028110.D
 Acq On : 14 Apr 2022 16:24
 Operator : JC/MD
 Sample : N2403-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-49

Quant Time: Apr 15 03:15:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	241496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	411532	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	386308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	177911	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	152683	46.602	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.200%
35) Dibromofluoromethane	5.391	113	134709	49.954	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.900%
50) Toluene-d8	8.653	98	498307	49.547	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	166408	44.314	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.620%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.953	59	632132	831.193	ug/l	100
16) Acetone	2.380	43	6804	4.984	ug/l	93
19) Methyl tert-butyl Ether	3.117	73	2623927	293.106	ug/l	100
20) Methylene Chloride	2.788	84	2148	0.790	ug/l #	85
22) Diisopropyl ether	3.763	45	7676	0.848	ug/l	92
25) 2-Butanone	4.580	43	3098	1.313	ug/l #	84
40) Benzene	6.044	78	64200	5.364	ug/l	98
67) Ethyl Benzene	10.195	91	11949	0.805	ug/l	98
68) m/p-Xylenes	10.305	106	8233	1.328	ug/l	99
69) o-Xylene	10.646	106	2000	0.332	ug/l	94
73) Isopropylbenzene	10.963	105	2935	0.232	ug/l	98
80) 1,3,5-Trimethylbenzene	11.457	105	2381	0.218	ug/l	95
84) 1,2,4-Trimethylbenzene	11.756	105	9434	0.883	ug/l	99
95) Naphthalene	13.780	128	10363	0.883	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041422\
Data File : VX028110.D
Acq On : 14 Apr 2022 16:24
Operator : JC/MD
Sample : N2403-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-49

Quant Time: Apr 15 03:15:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

