

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041422\
 Data File : VX028111.D
 Acq On : 14 Apr 2022 16:48
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 15 03:16:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	217847	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377208	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358861	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165014	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	139071	47.055	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.120%
35) Dibromofluoromethane	5.391	113	123754	50.068	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.140%
50) Toluene-d8	8.653	98	456395	49.509	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.020%
62) 4-Bromofluorobenzene	11.085	95	155656	45.223	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.440%

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041422\
Data File : VX028111.D
Acq On : 14 Apr 2022 16:48
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 15 03:16:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

