



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041422\
 Data File : VX028094.D
 Acq On : 14 Apr 2022 09:08
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Title : SW846 8260
 Last Update : Tue Apr 12 22:07:23 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.5	9793	PASS
75	95	30	60	53.5	24429	PASS
95	95	100	100	100.0	45643	PASS
96	95	5	9	6.2	2815	PASS
173	174	0.00	2	0.6	220	PASS
174	95	50	100	78.2	35683	PASS
175	174	5	9	8.3	2965	PASS
176	174	95	101	98.9	35277	PASS
177	176	5	9	6.6	2345	PASS