

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041423\
 Data File : VX035061.D
 Acq On : 13 Apr 2023 14:04
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 14 05:41:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	269104	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	452019	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	400763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	184540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	178592	43.394	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.780%
35) Dibromofluoromethane	5.385	113	139366	46.615	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.220%
50) Toluene-d8	8.647	98	535545	48.530	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.060%
62) 4-Bromofluorobenzene	11.079	95	202027	47.309	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.620%
Target Compounds						
94) Hexachlorobutadiene	13.725	225	1012	0.666	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041423\
Data File : VX035061.D
Acq On : 13 Apr 2023 14:04
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 14 05:41:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

