


BFB

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041423\
Data File : VX035053.D
Acq On : 13 Apr 2023 08:26
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Title : SW846 8260
Last Update : Sat Apr 01 01:50:18 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	18.0	7638	PASS
75	95	30	60	51.2	21731	PASS
95	95	100	100	100.0	42448	PASS
96	95	5	9	7.0	2963	PASS
173	174	0.00	2	1.3	386	PASS
174	95	50	100	71.0	30125	PASS
175	174	5	9	7.2	2175	PASS
176	174	95	101	95.9	28904	PASS
177	176	5	9	6.4	1846	PASS