

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045766.D
 Acq On : 14 Apr 2025 14:25
 Operator : JC/MD
 Sample : Q1799-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-216-PPR

Quant Time: Apr 15 01:28:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.550	168	66101	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	118297	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49980	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65656	54.314	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.620%	
35) Dibromofluoromethane	5.379	113	47036	51.196	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.400%	
50) Toluene-d8	8.647	98	161143	50.251	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.500%	
62) 4-Bromofluorobenzene	11.079	95	59188	50.671	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 101.340%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.294	50	318	0.313	ug/l #	42
5) Bromomethane	1.599	94	107	0.243	ug/l #	3
8) Diethyl Ether	2.136	74	222	0.477	ug/l #	4
11) Tert butyl alcohol	2.959	59	6819	41.975	ug/l #	92
13) Acrolein	2.233	56	5065	22.641	ug/l	94
16) Acetone	2.379	43	299826	604.033	ug/l	99
18) Methyl Acetate	2.544	43	6654	5.822	ug/l #	54
20) Methylene Chloride	2.788	84	199	0.214	ug/l #	84
22) Diisopropyl ether	3.763	45	1268	0.428	ug/l #	75
23) Vinyl Acetate	3.757	43	750	0.293	ug/l #	74
25) 2-Butanone	4.580	43	6377	8.777	ug/l	96
29) Tetrahydrofuran	5.050	42	102	0.217	ug/l #	35
31) Cyclohexane	5.537	56	2016	1.360	ug/l #	50
36) 1,1-Dichloropropene	5.550	75	6632	5.346	ug/l #	50
38) Carbon Tetrachloride	5.543	117	7555	5.635	ug/l #	17
51) 4-Methyl-2-Pentanone	8.573	43	12990	8.270	ug/l	100
52) Toluene	8.720	92	11438	4.989	ug/l	96
53) t-1,3-Dichloropropene	8.726	75	24	2.306	ug/l #	1
59) 2-Hexanone	9.463	43	580	0.499	ug/l	87
67) Ethyl Benzene	10.195	91	1436	0.317	ug/l #	80
68) m/p-Xylenes	10.299	106	1895	1.151	ug/l	94
69) o-Xylene	10.646	106	1317	0.812	ug/l	98
76) 1,2,3-Trichloropropane	11.079	75	34863	28.597	ug/l #	37
80) 1,3,5-Trimethylbenzene	11.451	105	1100	0.332	ug/l	91
81) trans-1,4-Dichloro-2-b...	11.079	75	34863	97.053	ug/l #	1
84) 1,2,4-Trimethylbenzene	11.756	105	3297	0.992	ug/l	97
89) n-Butylbenzene	12.329	91	669	0.232	ug/l #	60
90) Hexachloroethane	12.609	117	246	0.431	ug/l #	19
92) 1,2-Dibromo-3-Chloropr...	12.969	75	66	0.224	ug/l #	9
95) Naphthalene	13.774	128	21153	6.100	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045766.D
Acq On : 14 Apr 2025 14:25
Operator : JC/MD
Sample : Q1799-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR

Quant Time: Apr 15 01:28:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion:168 Resp: 66101
Ion Ratio Lower Upper
168 100
99 72.2 52.3 78.5



#3
Chloromethane
Concen: 0.313 ug/l
RT: 1.294 min Scan# 34
Delta R.T. -0.012 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 50 Resp: 318
Ion Ratio Lower Upper
50 100
52 0.0 25.9 38.9#









#5
Bromomethane
Concen: 0.243 ug/l
RT: 1.599 min Scan# 84
Delta R.T. 0.006 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion: 94 Resp: 107
Ion Ratio Lower Upper
94 100
96 0.0 74.8 112.2#



#8
Diethyl Ether
Concen: 0.477 ug/l
RT: 2.136 min Scan# 172
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 74 Resp: 222
Ion Ratio Lower Upper
74 100
45 205.4 53.0 159.1#









#11
Tert butyl alcohol
Concen: 41.975 ug/l
RT: 2.959 min Scan# 307
Delta R.T. -0.019 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion: 59 Resp: 6819
Ion Ratio Lower Upper
59 100
57 7.5 8.6 12.8#



#13
Acrolein
Concen: 22.641 ug/l
RT: 2.233 min Scan# 188
Delta R.T. -0.006 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 56 Resp: 5065
Ion Ratio Lower Upper
56 100
55 75.5 56.2 84.2









#16
Acetone
Concen: 604.033 ug/l
RT: 2.379 min Scan# 213
Delta R.T. -0.006 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion: 43 Resp: 299826
Ion Ratio Lower Upper
43 100
58 27.3 22.3 33.5



#18
Methyl Acetate
Concen: 5.822 ug/l
RT: 2.544 min Scan# 239
Delta R.T. -0.159 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 43 Resp: 6654
Ion Ratio Lower Upper
43 100
74 0.0 17.5 26.3#









#20
Methylene Chloride
Concen: 0.214 ug/l
RT: 2.788 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion:	84	Resp:	199
Ion	Ratio	Lower	Upper
84	100		
49	134.6	107.5	161.3
51	0.0	33.0	49.6#
86	76.7	50.1	75.1#



#22
Diisopropyl ether
Concen: 0.428 ug/l
RT: 3.763 min Scan# 439
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion:	45	Resp:	1268
Ion	Ratio	Lower	Upper
45	100		
43	48.0	59.3	88.9#
87	17.1	20.4	30.6#
59	14.6	9.0	13.6#









#23
 Vinyl Acetate
 Concen: 0.293 ug/l
 RT: 3.757 min Scan# 41
 Delta R.T. 0.036 min
 Lab File: VX045766.D
 Acq: 14 Apr 2025 14:25




Instrument :
 MSVOA_X
 ClientSampleId :
 FRAC-TANK-216-PPR


Tgt Ion: 43 Resp: 750
 Ion Ratio Lower Upper
 43 100
 86 0.0 7.5 11.3#



#25
 2-Butanone
 Concen: 8.777 ug/l
 RT: 4.580 min Scan# 573
 Delta R.T. 0.024 min
 Lab File: VX045766.D
 Acq: 14 Apr 2025 14:25

Tgt Ion: 43 Resp: 6377
 Ion Ratio Lower Upper
 43 100
 72 21.3 18.6 28.0





#29

Tetrahydrofuran

Concen: 0.217 ug/l

RT: 5.050 min Scan# 61

Delta R.T. 0.049 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :

FRAC-TANK-216-PPR





Tgt Ion: 42 Resp: 102

Ion Ratio Lower Upper

42 100

72 0.0 33.4 50.0#

71 0.0 31.4 47.0#



#31

Cyclohexane

Concen: 1.360 ug/l

RT: 5.537 min Scan# 730

Delta R.T. 0.073 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25




Tgt Ion: 56 Resp: 2016


Ion Ratio Lower Upper

56 100

69 112.1 23.6 35.4#

84 71.6 67.7 101.5





#33

1,2-Dichloroethane-d4

Concen: 54.314 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :





FRAC-TANK-216-PPR

Tgt Ion: 65 Resp: 65656

Ion Ratio Lower Upper

65 100

67 48.7 0.0 99.0



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.757 min Scan# 930

Delta R.T. -0.000 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25




Tgt Ion: 114 Resp: 129492


Ion Ratio Lower Upper

114 100

63 24.7 0.0 46.8

88 18.4 0.0 35.4









#35
Dibromofluoromethane
Concen: 51.196 ug/l
RT: 5.379 min Scan# 704
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion:113 Resp: 47036
Ion Ratio Lower Upper
113 100
111 101.5 81.8 122.6
192 16.9 13.8 20.6



#36
1,1-Dichloropropene
Concen: 5.346 ug/l
RT: 5.550 min Scan# 732
Delta R.T. -0.134 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25


Tgt Ion: 75 Resp: 6632
Ion Ratio Lower Upper
75 100
110 8.2 16.9 50.6#
77 0.0 25.0 37.4#








#38
Carbon Tetrachloride
Concen: 5.635 ug/l
RT: 5.543 min Scan# 71
Delta R.T. -0.128 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR







Tgt Ion:117 Resp: 7555
Ion Ratio Lower Upper
117 100
119 5.9 76.9 115.3#
121 0.0 24.2 36.2#



#50
Toluene-d8
Concen: 50.251 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 98 Resp: 161143
Ion Ratio Lower Upper
98 100
100 64.1 52.2 78.4









#51
4-Methyl-2-Pentanone
Concen: 8.270 ug/l
RT: 8.573 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion: 43 Resp: 12990
Ion Ratio Lower Upper
43 100
58 35.4 28.5 42.7



#52
Toluene
Concen: 4.989 ug/l
RT: 8.720 min Scan# 1252
Delta R.T. 0.006 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 92 Resp: 11438
Ion Ratio Lower Upper
92 100
91 179.5 138.7 208.1





#53

t-1,3-Dichloropropene

Concen: 2.306 ug/l

RT: 8.726 min Scan# 11

Delta R.T. -0.250 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :





FRAC-TANK-216-PPR

Tgt Ion: 75 Resp: 24

Ion Ratio Lower Upper

75 100

77 178.8 25.4 38.0#



#59

2-Hexanone

Concen: 0.499 ug/l

RT: 9.463 min Scan# 1374

Delta R.T. 0.036 min

Lab File: VX045766.D




Acq: 14 Apr 2025 14:25


Tgt Ion: 43 Resp: 580

Ion Ratio Lower Upper

43 100

58 40.5 24.9 74.6





#62

4-Bromofluorobenzene

Concen: 50.671 ug/l

RT: 11.079 min Scan# 1

Delta R.T. -0.000 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :

FRAC-TANK-216-PPR





Tgt Ion: 95 Resp: 59188

Ion Ratio Lower Upper

95 100

174 67.2 0.0 135.8

176 65.6 0.0 131.4



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.049 min Scan# 1470

Delta R.T. -0.000 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25




Tgt Ion: 117 Resp: 118297


Ion Ratio Lower Upper

117 100

82 64.1 49.2 73.8

119 33.1 25.1 37.7





#67

Ethyl Benzene

Concen: 0.317 ug/l

RT: 10.195 min Scan# 1493

Delta R.T. 0.006 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :





FRAC-TANK-216-PPR

Tgt Ion: 91 Resp: 1436

Ion Ratio Lower Upper

91 100

106 18.9 23.7 35.5#



#68

m/p-Xylenes

Concen: 1.151 ug/l

RT: 10.299 min Scan# 1511

Delta R.T. -0.000 min

Lab File: VX045766.D




Acq: 14 Apr 2025 14:25


Tgt Ion:106 Resp: 1895

Ion Ratio Lower Upper

106 100

91 206.8 172.6 258.8





#69

o-Xylene

Concen: 0.812 ug/l

RT: 10.646 min Scan# 1567

Delta R.T. 0.006 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :





FRAC-TANK-216-PPR

Tgt Ion:106 Resp: 1317

Ion Ratio Lower Upper

106 100

91 230.4 113.6 340.6



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.018 min Scan# 1793

Delta R.T. -0.000 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25




Tgt Ion:152 Resp: 49980


Ion Ratio Lower Upper

152 100

115 66.9 46.9 140.7

150 155.9 0.0 349.4





#76

1,2,3-Trichloropropane

Concen: 28.597 ug/l

RT: 11.079 min Scan# 1

Delta R.T. -0.159 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :





FRAC-TANK-216-PPR

Tgt Ion: 75 Resp: 34863

Ion Ratio Lower Upper

75 100

77 1.2 20.0 59.9#



#80

1,3,5-Trimethylbenzene

Concen: 0.332 ug/l

RT: 11.451 min Scan# 1700

Delta R.T. -0.000 min

Lab File: VX045766.D




Acq: 14 Apr 2025 14:25


Tgt Ion: 105 Resp: 1100

Ion Ratio Lower Upper

105 100

120 51.8 23.0 69.0









#81
trans-1,4-Dichloro-2-butene
Concen: 97.053 ug/l
RT: 11.079 min Scan# 1
Delta R.T. 0.061 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25




Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR


Tgt Ion: 75 Resp: 34863
Ion Ratio Lower Upper
75 100
53 0.0 101.8 152.6#
89 0.0 37.2 55.8#



#84
1,2,4-Trimethylbenzene
Concen: 0.992 ug/l
RT: 11.756 min Scan# 1750
Delta R.T. 0.006 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 105 Resp: 3297
Ion Ratio Lower Upper
105 100
120 44.9 21.4 64.3





#89

n-Butylbenzene

Concen: 0.232 ug/l

RT: 12.329 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX045766.D

Acq: 14 Apr 2025 14:25

Instrument :

MSVOA_X

ClientSampleId :

FRAC-TANK-216-PPR





Tgt Ion: 91 Resp: 669

Ion Ratio Lower Upper

91 100

92 30.8 26.9 80.6

134 51.6 11.9 35.7#



#90

Hexachloroethane

Concen: 0.431 ug/l

RT: 12.609 min Scan# 1890

Delta R.T. 0.073 min

Lab File: VX045766.D




Acq: 14 Apr 2025 14:25


Tgt Ion: 117 Resp: 246

Ion Ratio Lower Upper

117 100

201 0.0 31.2 93.6#









#92
1,2-Dibromo-3-Chloropropane
Concen: 0.224 ug/l
RT: 12.969 min Scan# 1949
Delta R.T. 0.030 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Instrument :
MSVOA_X
ClientSampleId :
FRAC-TANK-216-PPR

Tgt Ion: 75 Resp: 66
Ion Ratio Lower Upper
75 100
155 0.0 34.9 104.7#
157 0.0 44.9 134.7#



#95
Naphthalene
Concen: 6.100 ug/l
RT: 13.774 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VX045766.D
Acq: 14 Apr 2025 14:25

Tgt Ion: 128 Resp: 21153
Ion Ratio Lower Upper
128 100
127 13.3 10.3 15.5
129 11.0 8.6 12.8

