

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045757.D
 Acq On : 14 Apr 2025 10:41
 Operator : JC/MD
 Sample : VX0414WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0414WBL01

Quant Time: Apr 15 01:26:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129595	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119598	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50670	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65292	54.566	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.140%
35) Dibromofluoromethane	5.379	113	47382	51.531	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.060%
50) Toluene-d8	8.647	98	159858	49.810	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.620%
62) 4-Bromofluorobenzene	11.079	95	59909	51.248	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	102.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045757.D
Acq On : 14 Apr 2025 10:41
Operator : JC/MD
Sample : VX0414WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0414WBL01

Quant Time: Apr 15 01:26:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

