

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045760.D
 Acq On : 14 Apr 2025 11:55
 Operator : JC/MD
 Sample : Q1790-01
 Misc : 6.98g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S-872-G1-SO-20.0-20.5-041025

Quant Time: Apr 15 01:27:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68749	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	130423	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	115868	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48463	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	65244	51.895	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	103.780%	
35) Dibromofluoromethane	5.379	113	45597	49.275	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	98.560%	
50) Toluene-d8	8.647	98	160774	49.778	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	99.560%	
62) 4-Bromofluorobenzene	11.079	95	57459	48.840	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	97.680%	
Target Compounds						Qvalue
20) Methylene Chloride	2.776	84	1069	1.106	ug/l #	75
27) cis-1,2-Dichloroethene	4.483	96	1549	1.508	ug/l	88
44) Trichloroethene	7.129	130	2250	2.481	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045760.D
Acq On : 14 Apr 2025 11:55
Operator : JC/MD
Sample : Q1790-01
Misc : 6.98g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S-872-G1-SO-20.0-20.5-041025

Quant Time: Apr 15 01:27:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

