

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045768.D
 Acq On : 14 Apr 2025 15:12
 Operator : JC/MD
 Sample : Q1771-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-6C

Quant Time: Apr 15 01:29:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68783	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135123	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	132172	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	56416	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68050	54.100	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.200%
35) Dibromofluoromethane	5.385	113	48552	50.644	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.280%
50) Toluene-d8	8.647	98	172726	51.618	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.240%
62) 4-Bromofluorobenzene	11.079	95	67752	55.586	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	111.180%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	27317	52.887	ug/l	100
18) Methyl Acetate	2.709	43	1825	1.535	ug/l #	85
20) Methylene Chloride	2.782	84	3124	3.231	ug/l	85
25) 2-Butanone	4.580	43	4252	5.624	ug/l #	84
43) Isopropyl Acetate	6.348	43	7783	3.148	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045768.D
Acq On : 14 Apr 2025 15:12
Operator : JC/MD
Sample : Q1771-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-6C

Quant Time: Apr 15 01:29:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

