

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045769.D
 Acq On : 14 Apr 2025 15:35
 Operator : JC/MD
 Sample : Q1771-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-7C

Quant Time: Apr 15 01:29:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	74310	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145933	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	137028	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72398	53.275	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125		Recovery	= 106.560%	
35) Dibromofluoromethane	5.379	113	52565	50.768	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124		Recovery	= 101.540%	
50) Toluene-d8	8.647	98	183264	50.710	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113		Recovery	= 101.420%	
62) 4-Bromofluorobenzene	11.079	95	70497	53.554	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121		Recovery	= 107.100%	
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	24107	43.201	ug/l	97
18) Methyl Acetate	2.703	43	1683	1.310	ug/l	95
20) Methylene Chloride	2.788	84	3113	2.980	ug/l	92
43) Isopropyl Acetate	6.348	43	8247	3.089	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045769.D
Acq On : 14 Apr 2025 15:35
Operator : JC/MD
Sample : Q1771-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-7C

Quant Time: Apr 15 01:29:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

