

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045771.D
 Acq On : 14 Apr 2025 16:21
 Operator : JC/MD
 Sample : Q1779-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-22

Quant Time: Apr 15 01:29:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Amit Patel 04/15/2025
 Supervised By :Mahesh Dadoda 04/15/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	68866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	135348	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	125600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50121	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68192	54.147	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 108.300%			
35) Dibromofluoromethane	5.379	113	48841	50.860	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 101.720%			
50) Toluene-d8	8.647	98	172588	51.491	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 102.980%			
62) 4-Bromofluorobenzene	11.079	95	62388	51.100	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 102.200%			
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	24129	46.659	ug/l	96
18) Methyl Acetate	2.703	43	1895	1.592	ug/l	96
20) Methylene Chloride	2.782	84	3368	3.479	ug/l	94
25) 2-Butanone	4.587	43	4075m	5.383	ug/l	
43) Isopropyl Acetate	6.342	43	8475	3.422	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045771.D
Acq On : 14 Apr 2025 16:21
Operator : JC/MD
Sample : Q1779-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-22

Quant Time: Apr 15 01:29:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Amit Patel 04/15/2025
Supervised By :Mahesh Dadoda 04/15/2025

