

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045772.D
 Acq On : 14 Apr 2025 16:44
 Operator : JC/MD
 Sample : Q1779-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-23

Quant Time: Apr 15 01:29:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	65348	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	128538	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	122013	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	49069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64932	54.334	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.660%	
35) Dibromofluoromethane	5.379	113	46869	51.393	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.780%	
50) Toluene-d8	8.647	98	165038	51.847	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.700%	
62) 4-Bromofluorobenzene	11.079	95	60434	52.122	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 104.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	42025	85.640	ug/l	97
18) Methyl Acetate	2.709	43	2104	1.862	ug/l	99
20) Methylene Chloride	2.788	84	3375	3.674	ug/l #	89
25) 2-Butanone	4.574	43	8048	11.204	ug/l	98
43) Isopropyl Acetate	6.342	43	8764	3.726	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045772.D
Acq On : 14 Apr 2025 16:44
Operator : JC/MD
Sample : Q1779-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-23

Quant Time: Apr 15 01:29:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

