

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041425\  
 Data File : VX045775.D  
 Acq On : 14 Apr 2025 17:54  
 Operator : JC/MD  
 Sample : Q1781-12  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 WC-11

Quant Time: Apr 15 01:30:38 2025  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Apr 02 03:11:43 2025  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|--------|----------|
| -----                       |        |       |          |          |        |          |
| Internal Standards          |        |       |          |          |        |          |
| 1) Pentafluorobenzene       | 5.544  | 168   | 70031    | 50.000   | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 136080   | 50.000   | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.049 | 117   | 131577   | 50.000   | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.018 | 152   | 54629    | 50.000   | ug/l   | 0.00     |
| System Monitoring Compounds |        |       |          |          |        |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 70575    | 55.107   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 74 - 125 | Recovery | =      | 110.220% |
| 35) Dibromofluoromethane    | 5.379  | 113   | 49949    | 51.734   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =      | 103.460% |
| 50) Toluene-d8              | 8.647  | 98    | 176330   | 52.324   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 86 - 113 | Recovery | =      | 104.640% |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 65990    | 53.760   | ug/l   | 0.00     |
| Spiked Amount               | 50.000 | Range | 77 - 121 | Recovery | =      | 107.520% |
| Target Compounds            |        |       |          |          |        |          |
|                             |        |       |          |          | Qvalue |          |
| 16) Acetone                 | 2.380  | 43    | 25150    | 47.824   | ug/l   | 97       |
| 18) Methyl Acetate          | 2.703  | 43    | 2171     | 1.793    | ug/l # | 85       |
| 20) Methylene Chloride      | 2.782  | 84    | 3484     | 3.539    | ug/l   | 92       |
| 43) Isopropyl Acetate       | 6.354  | 43    | 8475     | 3.404    | ug/l # | 94       |
| -----                       |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041425\  
Data File : VX045775.D  
Acq On : 14 Apr 2025 17:54  
Operator : JC/MD  
Sample : Q1781-12  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
WC-11

Quant Time: Apr 15 01:30:38 2025  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X040225W.M  
Quant Title : SW846 8260  
QLast Update : Wed Apr 02 03:11:43 2025  
Response via : Initial Calibration

