

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045779.D
 Acq On : 14 Apr 2025 19:26
 Operator : JC/MD
 Sample : Q1788-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-1

Quant Time: Apr 15 01:31:32 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	69486	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	138197	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	127892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53349	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70404	55.405	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125		Recovery	= 110.800%	
35) Dibromofluoromethane	5.379	113	50282	51.282	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124		Recovery	= 102.560%	
50) Toluene-d8	8.647	98	177114	51.752	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113		Recovery	= 103.500%	
62) 4-Bromofluorobenzene	11.079	95	63884	51.247	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121		Recovery	= 102.500%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	25431	48.738	ug/l	96
18) Methyl Acetate	2.703	43	2340	1.948	ug/l #	87
20) Methylene Chloride	2.788	84	3289	3.367	ug/l #	87
43) Isopropyl Acetate	6.342	43	7343	2.904	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045779.D
Acq On : 14 Apr 2025 19:26
Operator : JC/MD
Sample : Q1788-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-1

Quant Time: Apr 15 01:31:32 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

