

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045782.D
 Acq On : 14 Apr 2025 20:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 15 01:32:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.748	0.731	2.3	87	0.00
3 P	Chloromethane	0.768	0.691	10.0	86	0.00
4 C	Vinyl Chloride	0.703	0.641	8.8#	87	0.00
5 T	Bromomethane	0.333	0.285	14.4	84	0.00
6 T	Chloroethane	0.373	0.375	-0.5	92	0.00
7 T	Trichlorofluoromethane	1.048	1.056	-0.8	94	0.00
8 T	Diethyl Ether	0.352	0.344	2.3	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.620	-1.0	97	0.00
10 T	Methyl Iodide	0.766	0.756	1.3	92	0.00
11 T	Tert butyl alcohol	0.123	0.124	-0.8	95	0.00
12 CM	1,1-Dichloroethene	0.600	0.562	6.3#	90	0.00
13 T	Acrolein	0.169	0.152	10.1	88	0.00
14 T	Allyl chloride	1.139	1.119	1.8	93	0.00
15 T	Acrylonitrile	0.388	0.378	2.6	90	0.00
16 T	Acetone	0.375	0.359	4.3	92	0.00
17 T	Carbon Disulfide	1.483	1.151	22.4	72	0.00
18 T	Methyl Acetate	0.864	0.986	-14.1	112	0.00
19 T	Methyl tert-butyl Ether	2.100	2.122	-1.0	97	0.00
20 T	Methylene Chloride	0.703	0.659	6.3	90	0.00
21 T	trans-1,2-Dichloroethene	0.613	0.570	7.0	89	0.00
22 T	Diisopropyl ether	2.243	2.279	-1.6	97	0.00
23 T	Vinyl Acetate	1.938	1.946	-0.4	91	0.00
24 P	1,1-Dichloroethane	1.269	1.232	2.9	94	0.00
25 T	2-Butanone	0.550	0.545	0.9	90	0.00
26 T	2,2-Dichloropropane	0.820	0.825	-0.6	95	0.00
27 T	cis-1,2-Dichloroethene	0.747	0.719	3.7	93	0.00
28 T	Bromochloromethane	0.613	0.567	7.5	92	0.00
29 T	Tetrahydrofuran	0.356	0.343	3.7	90	0.00
30 C	Chloroform	1.306	1.309	-0.2#	97	0.00
31 T	Cyclohexane	1.122	1.050	6.4	91	0.00
32 T	1,1,1-Trichloroethane	1.105	1.111	-0.5	97	0.00
33 S	1,2-Dichloroethane-d4	0.914	0.895	2.1	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	97	0.00
35 S	Dibromofluoromethane	0.355	0.350	1.4	99	0.00
36 T	1,1-Dichloropropene	0.479	0.466	2.7	94	0.00
37 T	Ethyl Acetate	0.604	0.579	4.1	90	0.00
38 T	Carbon Tetrachloride	0.518	0.528	-1.9	96	0.00
39 T	Methylcyclohexane	0.587	0.569	3.1	89	0.00
40 TM	Benzene	1.463	1.408	3.8	93	0.00
41 T	Methacrylonitrile	0.318	0.338	-6.3	94	-0.01
42 TM	1,2-Dichloroethane	0.604	0.619	-2.5	97	0.00
43 T	Isopropyl Acetate	0.915	0.924	-1.0	92	0.00
44 TM	Trichloroethene	0.348	0.331	4.9	92	0.00
45 C	1,2-Dichloropropane	0.365	0.364	0.3#	94	0.00
46 T	Dibromomethane	0.280	0.278	0.7	93	0.00
47 T	Bromodichloromethane	0.560	0.562	-0.4	96	0.00
48 T	Methyl methacrylate	0.472	0.490	-3.8	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
 Data File : VX045782.D
 Acq On : 14 Apr 2025 20:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 15 01:32:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	92	0.00
50 S	Toluene-d8	1.238	1.192	3.7	96	0.00
51 T	4-Methyl-2-Pentanone	0.606	0.625	-3.1	92	0.00
52 CM	Toluene	0.885	0.870	1.7#	93	0.00
53 T	t-1,3-Dichloropropene	0.467	0.492	-5.4	94	0.00
54 T	cis-1,3-Dichloropropene	0.526	0.550	-4.6	94	0.00
55 T	1,1,2-Trichloroethane	0.353	0.353	0.0	96	0.00
56 T	Ethyl methacrylate	0.546	0.575	-5.3	93	0.00
57 T	1,3-Dichloropropane	0.614	0.623	-1.5	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.276	0.297	-7.6	97	0.00
59 T	2-Hexanone	0.449	0.462	-2.9	91	0.00
60 T	Dibromochloromethane	0.385	0.393	-2.1	93	0.00
61 T	1,2-Dibromoethane	0.357	0.358	-0.3	93	0.00
62 S	4-Bromofluorobenzene	0.451	0.460	-2.0	99	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	96	0.00
64 T	Tetrachloroethene	0.353	0.333	5.7	92	0.00
65 PM	Chlorobenzene	1.068	1.071	-0.3	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.371	0.369	0.5	95	0.00
67 C	Ethyl Benzene	1.914	1.935	-1.1#	92	0.00
68 T	m/p-Xylenes	0.696	0.698	-0.3	92	0.00
69 T	o-Xylene	0.686	0.700	-2.0	93	0.00
70 T	Styrene	1.132	1.179	-4.2	93	0.00
71 P	Bromoform	0.277	0.278	-0.4	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
73 T	Isopropylbenzene	4.005	3.925	2.0	93	0.00
74 T	N-amyl acetate	1.916	1.959	-2.2	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.407	1.299	7.7	93	0.00
76 T	1,2,3-Trichloropropane	1.220	1.144	6.2	93	0.00
77 T	Bromobenzene	0.925	0.904	2.3	96	0.00
78 T	n-propylbenzene	4.613	4.567	1.0	93	0.00
79 T	2-Chlorotoluene	2.949	2.835	3.9	95	0.00
80 T	1,3,5-Trimethylbenzene	3.312	3.255	1.7	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.359	0.343	4.5	92	0.00
82 T	4-Chlorotoluene	3.288	3.271	0.5	94	0.00
83 T	tert-Butylbenzene	3.279	3.228	1.6	94	0.00
84 T	1,2,4-Trimethylbenzene	3.324	3.329	-0.2	94	0.00
85 T	sec-Butylbenzene	4.027	4.064	-0.9	94	0.00
86 T	p-Isopropyltoluene	3.320	3.303	0.5	93	0.00
87 T	1,3-Dichlorobenzene	1.684	1.631	3.1	95	0.00
88 T	1,4-Dichlorobenzene	1.708	1.614	5.5	94	0.00
89 T	n-Butylbenzene	2.879	2.973	-3.3	94	0.00
90 T	Hexachloroethane	0.571	0.577	-1.1	96	0.00
91 T	1,2-Dichlorobenzene	1.675	1.610	3.9	95	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.294	0.296	-0.7	93	0.00
93 T	1,2,4-Trichlorobenzene	0.932	0.923	1.0	97	0.00
94 T	Hexachlorobutadiene	0.402	0.393	2.2	97	0.00
95 T	Naphthalene	3.469	3.535	-1.9	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041425\
Data File : VX045782.D
Acq On : 14 Apr 2025 20:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 15 01:32:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.976	0.954	2.3	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6