

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041519\
 Data File : VX008956.D
 Acq On : 15 Apr 2019 12:22
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 16 03:36:54 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 02:27:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	250594	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	393401	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	335404	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	141484	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	157715	44.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.32%	
35) Dibromofluoromethane	5.49	113	118015	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	
50) Toluene-d8	8.71	98	488675	48.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.06%	
62) 4-Bromofluorobenzene	11.13	95	159988	45.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041519\
Data File : VX008956.D
Acq On : 15 Apr 2019 12:22
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 16 03:36:54 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 02:27:32 2019
Response via : Initial Calibration

