

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041519\
 Data File : VX008963.D
 Acq On : 15 Apr 2019 16:23
 Operator : JC/SP
 Sample : K2385-12DL 200X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS040MW320-190410DL

Quant Time: Apr 16 04:10:11 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 02:27:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274245	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	437810	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	378363	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	162143	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	176290	45.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.22%	
35) Dibromofluoromethane	5.50	113	129102	45.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.90%	
50) Toluene-d8	8.71	98	536366	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
62) 4-Bromofluorobenzene	11.13	95	180844	45.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.44%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	1907	0.535	ug/l	91
64) Tetrachloroethene	9.34	164	7204	2.506	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041519\
Data File : VX008963.D
Acq On : 15 Apr 2019 16:23
Operator : JC/SP
Sample : K2385-12DL 200X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS040MW320-190410DL

Quant Time: Apr 16 04:10:11 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 02:27:32 2019
Response via : Initial Calibration

