

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041519\
 Data File : VX008967.D
 Acq On : 15 Apr 2019 17:56
 Operator : JC/SP
 Sample : K2387-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-0508-190410

Quant Time: Apr 16 04:34:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 02:27:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	274494	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	438493	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	380802	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	172552	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	176513	45.63	ug/l	0.00
Spiked Amount			Recovery	=	91.26%	
35) Dibromofluoromethane	5.50	113	130880	46.51	ug/l	0.00
Spiked Amount			Recovery	=	93.02%	
50) Toluene-d8	8.71	98	546315	48.68	ug/l	0.00
Spiked Amount			Recovery	=	97.36%	
62) 4-Bromofluorobenzene	11.13	95	186898	47.18	ug/l	0.00
Spiked Amount			Recovery	=	94.36%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	4386	2.186	ug/l	97
21) trans-1,2-Dichloroethene	3.17	96	2738	0.872	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	611398	171.512	ug/l	90
30) Chloroform	5.21	83	6828	1.218	ug/l	93
44) Trichloroethene	7.21	130	598105	182.461	ug/l	99
47) Bromodichloromethane	7.90	83	2494	0.589	ug/l	96
64) Tetrachloroethene	9.34	164	2042624	706.087	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041519\
Data File : VX008967.D
Acq On : 15 Apr 2019 17:56
Operator : JC/SP
Sample : K2387-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-0508-190410

Quant Time: Apr 16 04:34:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 02:27:32 2019
Response via : Initial Calibration

