

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041519\
 Data File : VX008968.D
 Acq On : 15 Apr 2019 18:19
 Operator : JC/SP
 Sample : K2387-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW132-190411

Quant Time: Apr 16 04:37:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 02:27:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	267612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	434242	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	365508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150923	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	176156	46.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.42%	
35) Dibromofluoromethane	5.50	113	129456	46.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.90%	
50) Toluene-d8	8.71	98	527048	47.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.84%	
62) 4-Bromofluorobenzene	11.13	95	172035	43.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.70%	
Target Compounds						
4) Vinyl Chloride	1.41	62	4759	1.292	ug/l	100
27) cis-1,2-Dichloroethene	4.60	96	31547	9.077	ug/l	89
44) Trichloroethene	7.21	130	1719	0.530	ug/l	86
64) Tetrachloroethene	9.34	164	3750	1.351	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041519\
Data File : VX008968.D
Acq On : 15 Apr 2019 18:19
Operator : JC/SP
Sample : K2387-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW132-190411

Quant Time: Apr 16 04:37:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 02:27:32 2019
Response via : Initial Calibration

