

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041519\
 Data File : VX008969.D
 Acq On : 15 Apr 2019 18:42
 Operator : JC/SP
 Sample : K2387-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW130-190411

Quant Time: Apr 16 04:39:37 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 02:27:32 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	270093	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	429788	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	364731	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	150182	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	173863	45.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.36%	
35) Dibromofluoromethane	5.50	113	128035	46.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.84%	
50) Toluene-d8	8.71	98	531188	48.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.58%	
62) 4-Bromofluorobenzene	11.13	95	173099	44.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.16%	
Target Compounds						
4) Vinyl Chloride	1.41	62	2322	0.625	ug/l	94
16) Acetone	2.44	43	4174	2.115	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	114893	32.755	ug/l	90
44) Trichloroethene	7.21	130	6869	2.138	ug/l	94
64) Tetrachloroethene	9.34	164	3963	1.430	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041519\
Data File : VX008969.D
Acq On : 15 Apr 2019 18:42
Operator : JC/SP
Sample : K2387-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW130-190411

Quant Time: Apr 16 04:39:37 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 02:27:32 2019
Response via : Initial Calibration

