

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041520\
 Data File : VX015684.D
 Acq On : 15 Apr 2020 15:38
 Operator : JC/SP
 Sample : L2233-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW214-200409

Quant Time: Apr 16 07:00:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1070815	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1679322	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1565833	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	865890	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	654614	43.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.68%	
35) Dibromofluoromethane	5.47	113	503479	46.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.80%	
50) Toluene-d8	8.70	98	1978974	50.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.14%	
62) 4-Bromofluorobenzene	11.12	95	843378	52.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.02%	
Target Compounds						
16) Acetone	2.41	43	24228	3.734	ug/l #	89
44) Trichloroethene	7.20	130	4988	0.404	ug/l	70
64) Tetrachloroethene	9.32	164	18024	1.525	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041520\
Data File : VX015684.D
Acq On : 15 Apr 2020 15:38
Operator : JC/SP
Sample : L2233-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW214-200409

Quant Time: Apr 16 07:00:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

