

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041520\
 Data File : VX015686.D
 Acq On : 15 Apr 2020 16:23
 Operator : JC/SP
 Sample : L2233-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST008MW153-200408

Manual Integrations
 APPROVED

MMDadoda
 4/16/2020 11:03:30 AM

Quant Time: Apr 16 07:06:04 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1087241	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	1686592	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1607356	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	892391	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	659012	42.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	85.96%	
35) Dibromofluoromethane	5.47	113	517364	47.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.98%	
50) Toluene-d8	8.70	98	2006551	50.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.10%	
62) 4-Bromofluorobenzene	11.12	95	848307	52.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.18%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.39	62	3740	0.334	ug/l #	86
17) Carbon Disulfide	2.55	76	14180	0.453	ug/l #	84
27) cis-1,2-Dichloroethene	4.57	96	364293	27.640	ug/l	89
44) Trichloroethene	7.19	130	16152	1.301	ug/l	82
64) Tetrachloroethene	9.32	164	8574	0.707	ug/l	90
68) m/p-Xylenes	10.34	106	4690	0.222	ug/l	86
87) 1,3-Dichlorobenzene	12.01	146	6761	0.242	ug/l	99
88) 1,4-Dichlorobenzene	12.08	146	8327m	0.293	ug/l	
89) n-Butylbenzene	12.37	91	13069	0.276	ug/l	94
93) 1,2,4-Trichlorobenzene	13.63	180	9567	0.443	ug/l	94
94) Hexachlorobutadiene	13.76	225	6519	0.634	ug/l	99
95) Naphthalene	13.82	128	16743	0.254	ug/l #	93
96) 1,2,3-Trichlorobenzene	14.00	180	8419	0.405	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041520\
Data File : VX015686.D
Acq On : 15 Apr 2020 16:23
Operator : JC/SP
Sample : L2233-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST008MW153-200408

Manual Integrations
APPROVED

MMDadoda
4/16/2020 11:03:30 AM

Quant Time: Apr 16 07:06:04 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

