

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041520\
 Data File : VX015688.D
 Acq On : 15 Apr 2020 17:08
 Operator : JC/SP
 Sample : L2233-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-DUP-0544-200409

Quant Time: Apr 16 13:30:51 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1072643	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1668221	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1591869	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	870503	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	653049	43.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.34%	
35) Dibromofluoromethane	5.47	113	505900	47.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.88%	
50) Toluene-d8	8.70	98	1949725	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	11.12	95	828683	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
Target Compounds						
16) Acetone	2.42	43	9130	1.405	ug/l #	87
44) Trichloroethene	7.20	130	7285	0.593	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041520\
Data File : VX015688.D
Acq On : 15 Apr 2020 17:08
Operator : JC/SP
Sample : L2233-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-DUP-0544-200409

Quant Time: Apr 16 13:30:51 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

