

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041520\
 Data File : VX015689.D
 Acq On : 15 Apr 2020 17:31
 Operator : JC/SP
 Sample : L2233-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW206-200408

Quant Time: Apr 16 07:16:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	1013680	50.00	ug/l	0.01
34) 1,4-Difluorobenzene	6.84	114	1576565	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1509173	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	797121	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	631459	44.17	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.34%	
35) Dibromofluoromethane	5.47	113	480064	47.64	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	95.28%	
50) Toluene-d8	8.70	98	1874415	50.52	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	101.04%	
62) 4-Bromofluorobenzene	11.12	95	779707	51.71	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.42%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.37	101	19990	2.189	ug/l	89
16) Acetone	2.42	43	11817	1.924	ug/l	99
27) cis-1,2-Dichloroethene	4.58	96	6319	0.514	ug/l	85
30) Chloroform	5.19	83	8190	0.408	ug/l	94
37) Ethyl Acetate	4.81	43	23808	1.259	ug/l #	92
44) Trichloroethene	7.19	130	47574	4.099	ug/l	83
64) Tetrachloroethene	9.32	164	123016	10.801	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041520\
Data File : VX015689.D
Acq On : 15 Apr 2020 17:31
Operator : JC/SP
Sample : L2233-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW206-200408

Quant Time: Apr 16 07:16:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

