

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041520\
 Data File : VX015696.D
 Acq On : 15 Apr 2020 20:09
 Operator : JC/SP
 Sample : L2233-19DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z3-DUP-0543-200408DL

Quant Time: Apr 16 07:43:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 27 09:35:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	1010345	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	1584465	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	1547040	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	889869	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	625151	43.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.74%	
35) Dibromofluoromethane	5.47	113	466383	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.10%	
50) Toluene-d8	8.70	98	1912907	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
62) 4-Bromofluorobenzene	11.12	95	799936	52.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.56%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.56	96	85928	7.016	ug/l	89
44) Trichloroethene	7.19	130	50486	4.329	ug/l	98
64) Tetrachloroethene	9.32	164	649603	55.640	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041520\
Data File : VX015696.D
Acq On : 15 Apr 2020 20:09
Operator : JC/SP
Sample : L2233-19DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z3-DUP-0543-200408DL

Quant Time: Apr 16 07:43:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X032720W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 27 09:35:51 2020
Response via : Initial Calibration

