

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041520\  
 Data File : VX015698.D  
 Acq On : 15 Apr 2020 20:54  
 Operator : JC/SP  
 Sample : L2233-18  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 26 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SS037MW304-200408

Quant Time: Apr 16 07:50:01 2020  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X032720W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Mar 27 09:35:51 2020  
 Response via : Initial Calibration

| Internal Standards           | R.T.  | QIon | Response | Conc    | Units   | Dev(Min) |
|------------------------------|-------|------|----------|---------|---------|----------|
| 1) Pentafluorobenzene        | 5.64  | 168  | 1041680  | 50.00   | ug/l    | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.83  | 114  | 1617398  | 50.00   | ug/l    | 0.00     |
| 63) Chlorobenzene-d5         | 10.10 | 117  | 1553823  | 50.00   | ug/l    | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.06 | 152  | 855604   | 50.00   | ug/l    | 0.00     |
| System Monitoring Compounds  |       |      |          |         |         |          |
| 33) 1,2-Dichloroethane-d4    | 6.03  | 65   | 634961   | 43.22   | ug/l    | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =       | 86.44%  |          |
| 35) Dibromofluoromethane     | 5.47  | 113  | 495043   | 47.88   | ug/l    | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =       | 95.76%  |          |
| 50) Toluene-d8               | 8.70  | 98   | 1940602  | 50.98   | ug/l    | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =       | 101.96% |          |
| 62) 4-Bromofluorobenzene     | 11.12 | 95   | 808091   | 52.24   | ug/l    | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =       | 104.48% |          |
| Target Compounds             |       |      |          |         |         |          |
| 21) trans-1,2-Dichloroethene | 3.14  | 96   | 9004     | 0.825   | ug/l #  | 83       |
| 27) cis-1,2-Dichloroethene   | 4.57  | 96   | 443097   | 35.090  | ug/l    | 89       |
| 44) Trichloroethene          | 7.19  | 130  | 251765   | 21.147  | ug/l    | 94       |
| 64) Tetrachloroethene        | 9.32  | 164  | 3387781  | 288.904 | ug/l    | 92       |
| 71) Bromoform                | 10.84 | 173  | 12597    | 1.118   | ug/l #  | 97       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041520\  
Data File : VX015698.D  
Acq On : 15 Apr 2020 20:54  
Operator : JC/SP  
Sample : L2233-18  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 26 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SS037MW304-200408

Quant Time: Apr 16 07:50:01 2020  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X032720W.M  
Quant Title : SW846 8260  
QLast Update : Fri Mar 27 09:35:51 2020  
Response via : Initial Calibration

