

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
 Data File : VX028135.D
 Acq On : 15 Apr 2022 16:00
 Operator : JC/MD
 Sample : N2440-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TT-RW9-MW01D2-GW-20220413

Quant Time: Apr 16 02:26:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	218152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	379706	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	358417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165407	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142123	48.021	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.040%
35) Dibromofluoromethane	5.385	113	123414	49.602	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.200%
50) Toluene-d8	8.653	98	454953	49.028	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.060%
62) 4-Bromofluorobenzene	11.079	95	151574	43.747	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
Data File : VX028135.D
Acq On : 15 Apr 2022 16:00
Operator : JC/MD
Sample : N2440-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TT-RW9-MW01D2-GW-20220413

Quant Time: Apr 16 02:26:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

