

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
 Data File : VX028138.D
 Acq On : 15 Apr 2022 17:09
 Operator : JC/MD
 Sample : N2394-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-1-(DUMPSTER-RIGHT)

Quant Time: Apr 16 02:26:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	215511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	377539	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	385752	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	169917	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	136506	46.688	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	93.380%
35) Dibromofluoromethane	5.385	113	121775	49.224	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.440%
50) Toluene-d8	8.653	98	469611	50.898	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.800%
62) 4-Bromofluorobenzene	11.085	95	159762	46.375	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	92.760%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	8052	6.610	ug/l	97
20) Methylene Chloride	2.788	84	3827	1.577	ug/l #	85
43) Isopropyl Acetate	6.342	43	20503	2.798	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
Data File : VX028138.D
Acq On : 15 Apr 2022 17:09
Operator : JC/MD
Sample : N2394-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-1-(DUMPSTER-RIGHT)

Quant Time: Apr 16 02:26:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

