

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
 Data File : VX028139.D
 Acq On : 15 Apr 2022 17:33
 Operator : JC/MD
 Sample : N2394-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-2-(DUMPSTER-RIGHT)

Quant Time: Apr 16 02:27:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	219208	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	379402	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	370175	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	154953	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	141932	47.725	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.460%
35) Dibromofluoromethane	5.391	113	123131	49.528	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.060%
50) Toluene-d8	8.653	98	466369	50.298	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.600%
62) 4-Bromofluorobenzene	11.079	95	149458	43.171	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	86.340%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	7867	6.349	ug/l	98
20) Methylene Chloride	2.788	84	8779	3.557	ug/l	94
43) Isopropyl Acetate	6.342	43	19152	2.601	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
Data File : VX028139.D
Acq On : 15 Apr 2022 17:33
Operator : JC/MD
Sample : N2394-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-2-(DUMPSTER-RIGHT)

Quant Time: Apr 16 02:27:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

