

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
 Data File : VX028140.D
 Acq On : 15 Apr 2022 17:56
 Operator : JC/MD
 Sample : N2394-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-3-SOIL-PILE

Quant Time: Apr 16 02:27:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	213521	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	373273	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	394844	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173125	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	138063	47.661	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.320%
35) Dibromofluoromethane	5.391	113	120701	49.347	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.700%
50) Toluene-d8	8.652	98	465865	51.069	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.140%
62) 4-Bromofluorobenzene	11.079	95	161432	47.395	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.800%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	7818	6.478	ug/l	92
20) Methylene Chloride	2.788	84	5752	2.393	ug/l #	87
43) Isopropyl Acetate	6.348	43	18360	2.534	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
Data File : VX028140.D
Acq On : 15 Apr 2022 17:56
Operator : JC/MD
Sample : N2394-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-3-SOIL-PILE

Quant Time: Apr 16 02:27:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

