

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
 Data File : VX028142.D
 Acq On : 15 Apr 2022 18:42
 Operator : JC/MD
 Sample : N2394-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-5-(DUMPSTER-LEFT)

Quant Time: Apr 16 02:27:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	215922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	381750	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401140	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	173771	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	141745	48.388	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.780%
35) Dibromofluoromethane	5.391	113	122925	49.141	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.280%
50) Toluene-d8	8.653	98	470810	50.465	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.920%
62) 4-Bromofluorobenzene	11.079	95	160031	45.941	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.880%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	7093	5.812	ug/l	100
20) Methylene Chloride	2.788	84	5262	2.165	ug/l	93
43) Isopropyl Acetate	6.348	43	18291	2.469	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
Data File : VX028142.D
Acq On : 15 Apr 2022 18:42
Operator : JC/MD
Sample : N2394-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-5-(DUMPSTER-LEFT)

Quant Time: Apr 16 02:27:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

