

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
 Data File : VX028145.D
 Acq On : 15 Apr 2022 19:52
 Operator : JC/MD
 Sample : N2394-16
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CS-8

Quant Time: Apr 16 02:28:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 12 22:07:23 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	226764	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	391829	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	416272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	181184	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	147037	47.794	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.580%
35) Dibromofluoromethane	5.391	113	129555	50.459	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.920%
50) Toluene-d8	8.653	98	490010	51.172	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.340%
62) 4-Bromofluorobenzene	11.079	95	168737	47.194	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.380%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	8506	6.636	ug/l	89
20) Methylene Chloride	2.788	84	5659	2.217	ug/l	94
43) Isopropyl Acetate	6.342	43	18651	2.452	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041522\
Data File : VX028145.D
Acq On : 15 Apr 2022 19:52
Operator : JC/MD
Sample : N2394-16
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CS-8

Quant Time: Apr 16 02:28:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041222W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 12 22:07:23 2022
Response via : Initial Calibration

