



Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX041522\  
Data File : VX028121.D  
Acq On : 15 Apr 2022 08:54  
Operator : JC/MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041222W.M  
Title : SW846 8260  
Last Update : Tue Apr 12 22:07:23 2022



Spectrum Information: Average of 11.073 to 11.085 min.

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 21.5  | 6978  | PASS      |
| 75     | 95      | 30     | 60     | 57.2  | 18531 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 32400 | PASS      |
| 96     | 95      | 5      | 9      | 6.7   | 2180  | PASS      |
| 173    | 174     | 0.00   | 2      | 0.9   | 250   | PASS      |
| 174    | 95      | 50     | 100    | 86.5  | 28029 | PASS      |
| 175    | 174     | 5      | 9      | 7.4   | 2071  | PASS      |
| 176    | 174     | 95     | 101    | 95.2  | 26691 | PASS      |
| 177    | 176     | 5      | 9      | 6.9   | 1837  | PASS      |