

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041523\
 Data File : VX035080.D
 Acq On : 14 Apr 2023 13:59
 Operator : JC/MD
 Sample : 02351-01
 Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1691

Quant Time: Apr 17 05:09:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	287511	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	482935	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	423476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	191143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	185815	42.259	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	84.520%		
35) Dibromofluoromethane	5.385	113	135464	42.409	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	84.820%		
50) Toluene-d8	8.647	98	582762	49.428	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	98.860%		
62) 4-Bromofluorobenzene	11.085	95	220422	48.312	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.620%		
Target Compounds						Qvalue
39) Methylcyclohexane	7.373	83	15476	3.080	ug/l	91
52) Toluene	8.720	92	313947	35.444	ug/l	98
67) Ethyl Benzene	10.195	91	112012	7.018	ug/l	98
68) m/p-Xylenes	10.299	106	154500	25.340	ug/l	100
69) o-Xylene	10.640	106	88439	14.807	ug/l	100
73) Isopropylbenzene	10.964	105	65123	4.651	ug/l	99
78) n-propylbenzene	11.305	91	267967	16.567	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	286903	24.508	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	1111023	93.325	ug/l	98
85) sec-Butylbenzene	11.890	105	148153	11.078	ug/l	87
86) p-Isopropyltoluene	12.012	119	103375	9.270	ug/l	99
89) n-Butylbenzene	12.335	91	364319	37.393	ug/l #	53
95) Naphthalene	13.780	128	233344	18.350	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041523\
Data File : VX035080.D
Acq On : 14 Apr 2023 13:59
Operator : JC/MD
Sample : 02351-01
Misc : 1.02g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1691

Quant Time: Apr 17 05:09:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

