

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041523\
 Data File : VX035093.D
 Acq On : 14 Apr 2023 19:04
 Operator : JC/MD
 Sample : 02353-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BUR-1262

Quant Time: Apr 17 08:59:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	261256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	436810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	403740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	198845	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	171775	42.992	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.980%
35) Dibromofluoromethane	5.385	113	132168	45.746	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.500%
50) Toluene-d8	8.647	98	525800	49.306	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.079	95	204950	49.664	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.320%
Target Compounds						
					Qvalue	
16) Acetone	2.398	43	13135	8.092	ug/l	95
40) Benzene	6.044	78	2927	0.238	ug/l	99
52) Toluene	8.720	92	3934	0.491	ug/l	93
68) m/p-Xylenes	10.305	106	1576	0.271	ug/l	90
84) 1,2,4-Trimethylbenzene	11.750	105	4346	0.351	ug/l	86
95) Naphthalene	13.780	128	13563	1.025	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041523\
Data File : VX035093.D
Acq On : 14 Apr 2023 19:04
Operator : JC/MD
Sample : 02353-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BUR-1262

Quant Time: Apr 17 08:59:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

