

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041523\
 Data File : VX035094.D
 Acq On : 14 Apr 2023 19:27
 Operator : JC/MD
 Sample : 02308-01ME
 Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EO-01-041223

Quant Time: Apr 17 08:59:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
 Quant Title : SW846 8260
 QLast Update : Sat Apr 01 01:50:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	269853	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	451074	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	401704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	175766	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	170928	41.417	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	82.840%		
35) Dibromofluoromethane	5.385	113	127735	42.814	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	85.620%		
50) Toluene-d8	8.647	98	538813	48.929	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	97.860%		
62) 4-Bromofluorobenzene	11.085	95	193989	45.522	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.040%		
Target Compounds						
25) 2-Butanone	4.605	43	3963	1.625	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041523\
Data File : VX035094.D
Acq On : 14 Apr 2023 19:27
Operator : JC/MD
Sample : 02308-01ME
Misc : 5.10g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EO-01-041223

Quant Time: Apr 17 08:59:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X032923W.M
Quant Title : SW846 8260
QLast Update : Sat Apr 01 01:50:18 2023
Response via : Initial Calibration

