

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041524\
 Data File : VX041067.D
 Acq On : 15 Apr 2024 09:57
 Operator : JC/MD
 Sample : VX0415MBL01
 Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 16 02:49:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 16 02:47:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	324861	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	313613	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.018	152	168336	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	94471	40.885	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	81.780%
7) Chloroethane-d5	1.666	69	81567	43.284	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.560%
11) 1,1-Dichloroethene-d2	2.307	65	39437	39.977	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	79.960%
21) 2-Butanone-d5	4.446	46	135051	107.298	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	107.300%
24) Chloroform-d	5.050	84	180391	41.841	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.680%
26) 1,2-Dichloroethane-d4	5.952	65	121744	43.411	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.820%
32) Benzene-d6	5.971	84	353796	41.554	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.100%
36) 1,2-Dichloropropane-d6	7.306	67	108363	43.640	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.280%
41) Toluene-d8	8.647	98	333958	40.150	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.300%
43) trans-1,3-Dichloroprop...	8.946	79	51739	41.405	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.800%
47) 2-Hexanone-d5	9.385	63	94467	90.233	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.230%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	164051	44.600	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	89.200%
66) 1,2-Dichlorobenzene-d4	12.317	152	154082	45.630	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.260%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041524\
Data File : VX041067.D
Acq On : 15 Apr 2024 09:57
Operator : JC/MD
Sample : VX0415MBL01
Misc : 5.00g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 16 02:49:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 16 02:47:54 2024
Response via : Initial Calibration

