

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041524\
 Data File : VX041068.D
 Acq On : 15 Apr 2024 11:14
 Operator : JC/MD
 Sample : VX0415WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 16 02:49:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 16 02:47:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	285075	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	275702	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	147184	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	88929	43.858	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.720%
7) Chloroethane-d5	1.672	69	77129	46.641	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.280%
11) 1,1-Dichloroethene-d2	2.306	65	37775	43.636	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	87.280%
21) 2-Butanone-d5	4.452	46	128837	116.647	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	116.650%
24) Chloroform-d	5.056	84	172347	45.554	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.100%
26) 1,2-Dichloroethane-d4	5.958	65	115133	46.784	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.560%
32) Benzene-d6	5.970	84	336271	44.927	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.860%
36) 1,2-Dichloropropane-d6	7.305	67	101736	46.605	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	93.200%
41) Toluene-d8	8.647	98	314597	43.023	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.040%
43) trans-1,3-Dichloroprop...	8.951	79	47527	43.264	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.520%
47) 2-Hexanone-d5	9.384	63	87356	94.915	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.910%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	152032	47.016	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.040%
66) 1,2-Dichlorobenzene-d4	12.317	152	143767	48.693	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.380%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041524\
Data File : VX041068.D
Acq On : 15 Apr 2024 11:14
Operator : JC/MD
Sample : VX0415WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 16 02:49:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 16 02:47:54 2024
Response via : Initial Calibration

