

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041524\
 Data File : VX041074.D
 Acq On : 15 Apr 2024 13:35
 Operator : JC/MD
 Sample : P2098-03 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Apr 16 02:50:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 16 02:47:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.763	114	278275	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	267958	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	141250	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	81686	41.271	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	82.540%
7) Chloroethane-d5	1.666	69	72577	44.961	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.920%
11) 1,1-Dichloroethene-d2	2.306	65	34423	40.736	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	81.480%
21) 2-Butanone-d5	4.452	46	131175	121.666	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	121.670%
24) Chloroform-d	5.056	84	165883	44.917	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.840%
26) 1,2-Dichloroethane-d4	5.958	65	110326	45.926	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.860%
32) Benzene-d6	5.970	84	321967	44.259	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.520%
36) 1,2-Dichloropropane-d6	7.306	67	97134	45.783	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	91.560%
41) Toluene-d8	8.647	98	295340	41.556	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	83.120%
43) trans-1,3-Dichloroprop...	8.952	79	44995	42.143	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.280%
47) 2-Hexanone-d5	9.384	63	89283	99.812	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.810%
56) 1,1,2,2-Tetrachloroeth...	11.189	84	150584	47.914	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	95.820%
66) 1,2-Dichlorobenzene-d4	12.317	152	138917	49.027	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.060%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041524\
Data File : VX041074.D
Acq On : 15 Apr 2024 13:35
Operator : JC/MD
Sample : P2098-03 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Apr 16 02:50:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041024WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 16 02:47:54 2024
Response via : Initial Calibration

