

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041525\
 Data File : VX045786.D
 Acq On : 15 Apr 2025 09:55
 Operator : JC/MD
 Sample : VX0415WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0415WBL01

Quant Time: Apr 16 01:37:55 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64221	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124877	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48956	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63675	54.218	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.440%
35) Dibromofluoromethane	5.379	113	47222	53.298	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.600%
50) Toluene-d8	8.647	98	156579	50.632	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.260%
62) 4-Bromofluorobenzene	11.079	95	58980	52.359	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	104.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041525\
Data File : VX045786.D
Acq On : 15 Apr 2025 09:55
Operator : JC/MD
Sample : VX0415WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0415WBL01

Quant Time: Apr 16 01:37:55 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

