

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041525\
 Data File : VX045791.D
 Acq On : 15 Apr 2025 12:03
 Operator : JC/MD
 Sample : Q1798-02DL 250X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTE(COMP)-1799DL

Quant Time: Apr 16 01:39:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	71083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	136819	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	125654	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53400	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	69228	53.255	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.520%	
35) Dibromofluoromethane	5.379	113	50424	51.944	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.880%	
50) Toluene-d8	8.647	98	169597	50.055	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 100.100%	
62) 4-Bromofluorobenzene	11.079	95	63684	51.601	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.200%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	127252	238.396	ug/l	99
52) Toluene	8.720	92	926	0.382	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041525\
Data File : VX045791.D
Acq On : 15 Apr 2025 12:03
Operator : JC/MD
Sample : Q1798-02DL 250X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTE(COMP)-1799DL

Quant Time: Apr 16 01:39:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

