



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX041525\
Data File : VX045783.D
Acq On : 15 Apr 2025 08:28
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Title : SW846 8260
Last Update : Wed Apr 02 03:11:43 2025



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.5	7442	PASS
75	95	30	60	57.9	19150	PASS
95	95	100	100	100.0	33099	PASS
96	95	5	9	6.8	2265	PASS
173	174	0.00	2	0.7	156	PASS
174	95	50	100	66.0	21835	PASS
175	174	5	9	8.2	1782	PASS
176	174	95	101	99.8	21789	PASS
177	176	5	9	7.0	1530	PASS