

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX008983.D
 Acq On : 16 Apr 2019 13:32
 Operator : JC/SP
 Sample : K2387-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS042MW130-190411

Quant Time: Apr 17 05:28:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	262771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	416172	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	360685	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	152565	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168296	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.90%	
35) Dibromofluoromethane	5.50	113	124036	46.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.88%	
50) Toluene-d8	8.71	98	517664	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
62) 4-Bromofluorobenzene	11.13	95	173071	46.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.06%	
Target Compounds						
4) Vinyl Chloride	1.41	62	2175	0.601	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	114115	33.440	ug/l	90
44) Trichloroethene	7.21	130	6503	2.090	ug/l	97
64) Tetrachloroethene	9.33	164	2487	0.908	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX008983.D
Acq On : 16 Apr 2019 13:32
Operator : JC/SP
Sample : K2387-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS042MW130-190411

Quant Time: Apr 17 05:28:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

