

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX008985.D
 Acq On : 16 Apr 2019 14:18
 Operator : JC/SP
 Sample : K2387-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW145-190410

Quant Time: Apr 17 05:31:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	259278	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	415720	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	363833	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	157564	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	168546	46.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.26%	
35) Dibromofluoromethane	5.50	113	123310	46.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.44%	
50) Toluene-d8	8.71	98	520619	48.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.86%	
62) 4-Bromofluorobenzene	11.13	95	178051	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	2931	0.870	ug/l	91
44) Trichloroethene	7.22	130	1870	0.602	ug/l	74
64) Tetrachloroethene	9.34	164	5513	1.995	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX008985.D
Acq On : 16 Apr 2019 14:18
Operator : JC/SP
Sample : K2387-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW145-190410

Quant Time: Apr 17 05:31:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

