

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX008988.D
 Acq On : 16 Apr 2019 15:28
 Operator : JC/SP
 Sample : K2425-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190415

Quant Time: Apr 17 05:39:07 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	259203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	414728	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	367981	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158239	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	167175	45.76	ug/l	0.00
Spiked Amount			Recovery	=	91.52%	
35) Dibromofluoromethane	5.50	113	123395	46.36	ug/l	0.00
Spiked Amount			Recovery	=	92.72%	
50) Toluene-d8	8.71	98	518211	48.82	ug/l	0.00
Spiked Amount			Recovery	=	97.64%	
62) 4-Bromofluorobenzene	11.13	95	178192	47.56	ug/l	0.00
Spiked Amount			Recovery	=	95.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX008988.D
Acq On : 16 Apr 2019 15:28
Operator : JC/SP
Sample : K2425-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190415

Quant Time: Apr 17 05:39:07 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

