

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX041619\
 Data File : VX008995.D
 Acq On : 16 Apr 2019 18:11
 Operator : JC/SP
 Sample : K2387-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 KY062MW143-190410

Quant Time: Apr 17 05:59:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 16 09:35:21 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	269867	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	431629	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	379101	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	164118	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	174221	45.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.62%	
35) Dibromofluoromethane	5.50	113	126185	45.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.10%	
50) Toluene-d8	8.71	98	540958	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
62) 4-Bromofluorobenzene	11.13	95	184482	47.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.62%	
Target Compounds						
16) Acetone	2.45	43	3997	2.027	ug/l	97
27) cis-1,2-Dichloroethene	4.60	96	10352	2.954	ug/l	88
44) Trichloroethene	7.21	130	4204	1.303	ug/l	94
64) Tetrachloroethene	9.34	164	31087	10.794	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX041619\
Data File : VX008995.D
Acq On : 16 Apr 2019 18:11
Operator : JC/SP
Sample : K2387-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
KY062MW143-190410

Quant Time: Apr 17 05:59:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 16 09:35:21 2019
Response via : Initial Calibration

